

## Human adenovirus (ADV) antibody (IgM) ELISA kit

### **PRECISION**

**Intra-assay Precision (Precision within an assay): CV%<15%**

Three samples of known concentration were tested twenty times on one plate to assess.

**Inter-assay Precision (Precision between assays): CV%<15%**

Three samples of known concentration were tested in twenty assays to assess.

### **TYPICAL DATA**

Test parameter	specification	test result
Positive control	>1.0	1.543
Negative control	<0.2	0.105
Positive rate	10, Positive	100%
Negative rate	10, Negative	100%