

Dog anti-toxoplasmosis (TOXO) antibody (IgG) ELISA kit

PRECISION

Intra-assay Precision (Precision within an assay): CV%<15%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<15%

Three samples of known concentration were tested in twenty assays to assess.

TYPICAL DATA

Test parameter	specification	test result
Positive control	>1.0	1.538
Negative control	<0.1	0.046
Positive rate	10, Positive	100%
Negative rate	10, Negative	100%