

Certificate of Analysis

Product name: Human Tenascin C ELISA Kit

Catalog No.: MBS723481 **Lot No:** 20220826S **Exp:** Feb 26,2023

Manufacture date: Aug 26,2022 **Valid Period :** 6 months **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 6.9 7.9 Intra-assay CV%: 5.7 6.6

RECOVERY


Sample type	average% recovery	range
Serum	96	91-101%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	94-101
1: 4	Average % of Expected	93
	Range (%)	90-105
1: 8	Average % of Expected	89
	Range (%)	84-109

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	0.174
B	10	0.338
C	25	0.527
D	50	0.805
E	100	1.219
F	250	1.821



Date: 2022.08.26