

Certificate of Analysis

Product name: Rat Dopamine ELISA Kit

Cat No.: MBS725908 **Lot No:** 20210906C **Valid Period:** six months

Manufacturing date: Sep 06, 2021 **Exp:** Mar 06, 2022 **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 6.8 7.5 Intra-assay CV%: 5.7 7. 1

RECOVERY


Sample type	average% recovery	range
Serum	96	90-102%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	93-101
1: 4	Average % of Expected	93
	Range (%)	89-104
1: 8	Average % of Expected	89
	Range (%)	86-109

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	1.998
B	5	1.406
C	10	1.017
D	25	0.567
E	50	0.359
F	100	0.23



Date: 2021.09.06