

# Certificate of Analysis

**Product name:** Mouse Follicle Stimulating Hormone ELISA Kit

**Catalog No.:** MBS727159

**Lot No:** 20230609C

**Exp:** Dec 09,2023

**Manufacture date:** Jun 06,2023

**Valid Period :** 6 months

**Storage:** 2-8° C

## PRECISION

Inter-assay CV%: 6.2 7.9

Intra-assay CV%: 5.5 6.2

## RECOVERY

Sample type

average% recovery

range

Serum

98


93-103%

## LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	95-100
1: 4	Average % of Expected	93
	Range (%)	90-106
1: 8	Average % of Expected	92
	Range (%)	89-108

## STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	2.125
B	2.5	1.485
C	5	1.131
D	10	0.791
E	25	0.456
F	50	0.287



Date: 2023.06.09