

Certificate of Analysis

Human HMGCR(3-hydroxy-3-methylglutaryl-coenzyme A reductase) ELISA Kit

Cat.No.: **MBS760848**

Size: 96T


Lot:H1176E081

Principle: This ELISA kit uses sandwich-ELISA as the method

Product expiration Date: 2020/02/06

This product has been tested by Quality Control Department and meets performance specifications mentioned in the manual.

Concentration(ng/ml)	OD-1	OD-2	Average	OD-blank
0	0.071	0.073	0.072	0.000
0.156	0.162	0.170	0.166	0.094
0.312	0.292	0.306	0.299	0.227
0.625	0.503	0.479	0.491	0.419
1.25	0.772	0.794	0.783	0.711
2.5	1.248	1.312	1.28	1.208
5	1.793	1.749	1.771	1.699
10	2.149	2.257	2.203	2.131



Performance Characteristics

Intra-Assay CV: $\leq 8\%$

Inter-Assay CV: $\leq 10\%$

Date: 2019/08/06