

Quality Inspection Report

Product No.	MBS761262_QC	Batch No.	H0556B065
Product Name	Human NOS2/iNOS (Nitric oxide synthase, inducible) ELISA Kit	Inspection Date	6/16/2016
Size	96T		
Inspection Item	Standard requirements	Inspection Result	
Appearance	Complete outer packing and without damage		
	Correct labels		
	Correct specification in struction		
	Complete Reagent,cover sealed		
	Clear liquid,no sediment or flocs		
	Sealed plate		
	Complete components		
Standard Curve	The linear correlation coefficient should not be less than 0.9900		
Coefficient of Variation	Not larger than 10%		