
Certificate of Analysis

Human MUC-1(Mucin-1) ELISA Kit

Cat.No.: MBS762199

Size: 96T


Lot:H0406D055

Principle: This ELISA kit use Sandwich-ELISA as the method

Product expiration Date: 2018/11/16

This product has been tested by Quality Control Department and meets performance specifications mentioned in the manual.

Concentration(ng/ml)	OD-1	OD-2	Average	OD-blank
0	0.108	0.112	0.11	0.000
0.156	0.188	0.194	0.191	0.081
0.312	0.252	0.260	0.256	0.146
0.625	0.397	0.409	0.403	0.293
1.25	0.621	0.639	0.63	0.520
2.5	0.984	1.012	0.998	0.888
5	1.530	1.574	1.552	1.442
10	2.298	2.364	2.331	2.221



Performance Characteristics

Intra-Assay CV: $\leq 8\%$

Inter-Assay CV: $\leq 10\%$