


**Product Name: General Diacylglycerol ELISA Kit****Specificity:** Recombinant and natural General Diacylglycerol**Synonyms:** Diacylglycerol,;Diglyceride,;**Catalog No.: MBS2887354 Competitive****Standard Preparation**

Reconstitute the Standard with 1.0 mL of Sample Diluent. This reconstitution produces a stock solution of 40 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making serial dilutions (Making serial dilution in the wells directly is not permitted). The undiluted standard serves as the high standard (40 ng/mL). The Sample Diluent serves as the zero standard (0 ng/mL).

**Detection Range:** 0.625-40 ng/mL**Sensitivity:** less than 0.262 ng/mL**Performance Characteristics:** Intra-Assay CV: ≤6.5%


Inter-Assay CV: ≤8.9%

Spike Average Recovery: 96%

**Typical Data**

Results of a typical standard curve are provided for demonstration only and should not be used to obtain test results.

A standard curve must be run for each set of samples assayed.



**IMPORTANT:** After receiving the kit, please check the intactness of outer packaging, check whether the reagent bottles are broken, leaking or the amount of liquid is significantly reduced, etc. If the diluent is polluted such as flocculent bacterial colonies; the substrate is not transparent and the microtiter plate is leaking, etc; please contact the distributor. In rare cases, the crystals may appear in the substrate. Normally the crystals will dissolve at room temperature, this will not affect the testing results.